



TECHNOLEIA D-Dimer Control low

LOT K831C00

 2025-12-31

3015871- 031 14/02/2024

Ceveron 100 series barcode



Test test	Gerät instrument	Einheit unit	Richtwert (Vertrauensbereich) expected value (confidence range)	Std.
TECHNOLEIA D-Dimer	Ceveron 100 series	ng/mL	259 (194 - 324)	internal Ref. Std.
	Thrombolyzer	ng/mL	254 (191 - 318)	internal Ref. Std.